

CEREC – THE MOST IMPORTANT CLINICAL STUDIES

# SCIENTIFICALLY SECURE.



Mörmann ■ State of the Art of CAD/CAM Restorations



Vollkeramik auf einen Blick

AG Keramik



Die CEREC Computer Reconstruction

Mörmann / Brondum

Highly Esthetic CEREC Solutions

Dr. Rich Masek



INTERNATIONAL SYMPOSIUM ON COMPUTER RESTORATIONS



W. H. Mörmann

CEREC 3D DESIGN

W. H. Mörmann

Vollkeramische CAD-CAM  
Inlays und Teilkronen

# CONTENTS

Foreword and word of thanks .....	03
<b>1 Long-term performance of CEREC restorations.....</b>	<b>04</b>
1.1 Inlays/onlays.....	04
1.1.1 Long-term study of 2,328 chairside inlays/onlays .....	04
1.1.2 Eighteen-year study of 1,011 inlays/onlays .....	05
1.2 Veneers.....	06
1.3 Crowns.....	07
1.4 Comparison with other restoration types.....	08
1.4.1 Clinical comparison .....	08
1.4.2 Longevity and cost-effectiveness.....	09
1.4.3 Longevity and production costs.....	09
<b>2 Precision .....</b>	<b>10</b>
2.1 Image precision.....	10
2.1.1 Single tooth .....	10
2.1.2 Quadrant.....	10
2.2 Milling precision .....	11
2.2.1 Camera/milling unit.....	11
2.2.2 Marginal fit of restorations .....	11
<b>3 Marginal gap.....</b>	<b>12</b>
3.1 Adhesive interface.....	12
3.1.1 Materials .....	12
3.1.2 Marginal seal .....	13
3.1.3 Wear of the adhesive interface.....	13
3.2 Comparison of other restoration types.....	14
3.2.1 Enamel integrity .....	14
3.2.2 Margin quality .....	14
<b>4 Occlusal design.....</b>	<b>15</b>
4.1 Software .....	15
<b>5 Aesthetics .....</b>	<b>16</b>
5.1 Posterior teeth.....	16
5.2 Anterior teeth.....	17
5.2.1 Veneers .....	17
<b>6 Ceramic materials.....</b>	<b>18</b>
6.1 Strength/fracture toughness .....	19
6.2 Abrasion.....	20
CEREC publications .....	21

# FOREWORD AND WORD OF THANKS

## QUOD EST EST – WHAT IS, IS.

In the final analysis it all comes down to hard facts and evidence. This is precisely the purpose of the present compendium. Our aim is to summarize the latest clinical studies relating to CEREC so that you are in a position to interpret and evaluate the scientific findings.

CEREC certainly ranks as one of the most intensively scrutinized dental procedures – as evidenced in numerous clinical studies and a wide range of scientific publications.

For example, universities and scientifically oriented dental practices are continuously monitoring the survival rates of CEREC restorations (inlays, onlays, crowns and veneers) which were created and placed during a single appointment. The projected long-term survival rates are as high as 84.4 per cent after 18 years.

In terms of quality CEREC restorations are at least on a par with cast gold – and clearly superior to composite fillings and other laboratory-produced restorations.

Computer-aided dentistry has progressed enormously. The marginal gaps have reached laboratory standards. The design of the proximal contacts has become very reliable. The occlusal surfaces contained in the dental databases have been compiled by universities and renowned dental technicians. The CEREC system makes allowance for the patient's articulation and antagonists.

All that is needed is a good dentist – someone like you.

Over the past 20 years numerous persons have contributed to the further development of Professor Mörmann's original idea – i.e. to create high-quality ceramic restorations during a single appointment. This applies firstly to the members of the research teams at Siemens, Sirona, Vita Zahnfabrik, Ivoclar Vivadent, Merz, Zeiss and at numerous small and medium-sized enterprises. Secondly, more than 200 universities worldwide have conducted detailed research and made countless improvements – both large and small – to the CEREC procedure. Mention must also be made of the CEREC users, the CEREC instructors, the International Society for Computerized Dentistry and its national organizations. All these persons and organizations have played a pivotal role in ensuring that CEREC has become an integral part of modern dentistry. We would like to express our sincere thanks to all concerned. We would also like to thank the German Society for Computerized Dentistry for its expert help in the preparation of this compendium as well as the selection and interpretation of the scientific studies.

Your CEREC-Team

# 1 LONG-TERM PERFORMANCE OF CEREC RESTORATIONS


## 1.1 INLAYS/ONLAYS

### 1.1.1 LONG-TERM STUDY OF 2,328 CHAIRSIDE INLAYS/ONLAYS

This extensive study centred on 2,328 chairside CEREC inlays and onlays which had been fitted to a total of 794 patients in a dental practice. Between 1990 and 1997 the CEREC 1 system had been used; between 1997 and 1999 the CEREC 2 system was used.

Forty-four teeth were randomly selected and examined under a scanning electron microscope. The average margin width was  $236 \mu\text{m} \pm 96,8 \mu\text{m}$ .

The success rate after nine years was 95.5%. Only 35 restorations failed, due mainly to the extraction of the teeth. There was no correlation between failure and the size or location of the restorations.



Survival curve with regard to CP treatment

#### CONCLUSION

The long-term results (95.5% survival after nine years) are excellent, although CEREC 1 and CEREC 2 do not achieve today's level of clinical precision and the quality of the margins (created using macrofilled luting materials) does not conform to today's standards.

### 1.1.2 EIGHTEEN-YEAR STUDY OF 1,011 INLAYS/ONLAYS


This study centred on 1,011 CEREC inlays/onlays which had been fabricated for 299 patients between 1987 and 1990 using the CEREC 1 system. The majority of the restorations were made of VITA MK I ceramic; only a small number (22) were made of Dicor MGC. As from 1989 enamel etching (phosphoric acid) was deployed in combination with the dental adhesive Gluma. Glass ionomer cement was no longer used as the base layer. Areas close to the pulp were protected by means of a  $\text{CaOH}_2$  liner.

The follow-up criteria were as follows: margin quality, change in vitality, tooth anatomy, complications, and failures.

The findings were categorized according to the following parameters: restoration size, restoration location, initial tooth vitality, and the use of dentin adhesive.

During the 18-year observation period 86 of the 1,011 restorations were lost. Ceramic fractures were the main cause (38%).

According to the Kaplan Meier estimator, the probability of success after 18 years was extremely high (84.4%). Premolars perform slightly better than molars, and 2- and 3-surface inlays better than 1-surface inlays. There is a significance between non-vital teeth (50%) and vital teeth (88%). The application of a functional dentin adhesive increased the success rate by 10% to 90%.



Kaplan-Meier estimator: Use of dentin adhesive, n = 1.011


#### CONCLUSION

CEREC restorations (including those of a larger size) display outstanding longevity. In many cases defect-oriented restorations and careful adhesive bonding provide the basis for dispensing with full crowns.

# 1 LONG-TERM PERFORMANCE OF CEREC RESTORATIONS

## 1.2 VENEERS

The durability of laboratory-produced ceramic veneers has already been extensively researched. A group of CEREC veneers and partial anterior crowns was observed over a period of 9.5 years. These restorations had been produced on the CEREC 1 and CEREC 2 systems using VITA Mark II (mainly) and Ivoclar ProCad. 509 of the veneers had been bonded to natural teeth; 108 had been used to repair/replace existing PFM or gold-composite restorations. After 9.5 years the restorations attached to prosthetic elements had a success rate of 91%, while those placed on natural teeth showed a success rate of 94%.



Ceramic build-ups comprising up to 2/3 of the veneer length do not far worse.

### CONCLUSION

In terms of their longevity CEREC veneers do not differ from laboratory-produced veneers.

## 1.3 CROWNS

Following the introduction of CEREC 2 dentists were in a position to produce full crowns in addition to inlays and veneers.

In a further scientific study 208 CEREC crowns made of VITA Mark II were fitted to 136 patients using the adhesive bonding technique. Seventy of these crowns were placed on conventionally prepared teeth; 52 were placed on teeth with reduced stump preparations (low macroretention); and 86 crowns were placed on endodontically treated teeth. In this case the crowns included an additional post extending into the pulp cavity in order to achieve improved retention (endocrowns).

The main causes of failure were fractures, presumably due to inadequate dentin adhesion.

The “classic” crowns performed best of all (97.0% survival rate), followed by the “reduced” crowns (92.9%). The survival rate of the endocrowns was acceptable in the case of molars (87.1%) and relatively poor in the case of premolars (68.8%).

CEREC crowns also performed well in a study conducted in a dental practice. This study centred on 65 full crowns made of VITA Mark II which had been manually polished after the milling process and then bonded using dual-curing composite.

Three failures were observed in the period up to four years (two ceramic fractures, one debonding). The success rate according to Kaplan-Meier was 95.4%.

### CONCLUSION

CEREC crowns made of VITA Mark II and Ivoclar ProCad achieve success rates which are comparable to those PFM crowns.

# 1 LONG-TERM PERFORMANCE OF CEREC RESTORATIONS

## 1.4 COMPARISON WITH OTHER RESTORATION TYPES

### 1.4.1 CLINICAL COMPARISON

Long-term comparison of CEREC, laboratory-fabricated ceramic and gold inlays over a period of 15 years.

For the past 15 years 358 two- and three-surface inlays have been under observation at Graz University in Austria. The following restorations were placed on vital teeth: 93 gold inlays cemented with zinc phosphate cement (= control group); 71 adhesively bonded gold inlays; 94 laboratory-fabricated ceramic inlays (Dicor, Optec, Duceram, Hi-Ceram); and 51 CEREC inlays (VITA Mark I). In addition, a number of non-vital teeth were treated: gold/cement (5); gold/adhesive (14); laboratory-fabricated ceramic (22); and CEREC (8).

The restorations were assessed according to the following criteria: loss or complete fracture; partial fracture of the restoration, the tooth or the cement/adhesive bond; secondary caries; loss of tooth vitality. A Kaplan-Meier survival analysis was carried out for each group. In all groups inlays placed on non-vital teeth performed worse than inlays placed on vital teeth.

Initially the study included a group of indirect composite inlays. However, these were excluded prematurely on account of their very poor performance.

There was no significant statistical difference between the gold inlay groups and the CEREC inlays (success rate of approx. 93% after 15 years). The laboratory-fabricated ceramic inlays were clearly inferior (68%).

### CONCLUSION

In terms of longevity CEREC inlays are on a par with gold restorations. The laboratory-fabricated ceramic restorations performed worse.

The following ductile filling materials are used for posterior cavities: amalgam; glass ionomer and derivative products; and composites. In addition the following restoration types are available: gold inlays/onlays; composite inlays/onlays; laboratory-fabricated ceramic inlays/onlays; and CEREC inlays/onlays. Long-term studies have been carried out for each group.

These have revealed significant differences in longevity. The annual failure rate was determined for each restoration type. The ranking (from bad to good) is as follows:

7. Glass ionomer and derivative products (7.7%)
6. Amalgam (3.3%)
5. Composite fillings (2.2%)
4. Composite inlays/onlays (2.0%)
3. Ceramic inlays/onlays (1.6%)
2. Gold inlays/onlays (1.2%)
1. CEREC inlays/onlays (1.1%)


### CONCLUSION

The success rates of CEREC restorations are marginally better than those of gold inlays/onlays.

### 1.4.2 LONGEVITY AND COST-EFFECTIVENESS

In times of financial constraint it makes sense to evaluate the longevity and cost of dental restorations – not in isolation but in combination – in order to develop cost-effective restoration options for patients.

On the basis of billing data provided by a major German insurer the average fees and laboratory costs were determined for gold inlays [62], laboratory-fabricated ceramic inlays [87] and CEREC inlays [91]. A meta analysis was then performed of ten suitable long-term studies from the period 1994 to 2003. This provided the basis for determining the statistical longevity of the various inlay types.



### 1.4.3 LONGEVITY AND PRODUCTION COSTS

Due to their higher production costs and slightly lower survival probability, laboratory-fabricated ceramic inlays are the least cost-effective option. Gold inlays and CEREC inlays have similar success rates. However, given the higher laboratory costs of gold inlays, CEREC inlays emerge from this study as the most cost-effective restoration type.

### CONCLUSION

From an economic viewpoint CEREC inlays are preferable to all other inlay types.

# 2 PRECISION

## 2.1 IMAGE PRECISION

The precision of a milled CEREC restoration depends to a large extent on the quality of the data derived from the digital optical impression. The intraoral CEREC Bluecam has an innovative optical lens emitting blue light with a short wavelength.

### 2.1.1 SINGLE TOOTH

The scanning accuracy of CEREC Bluecam is approx. 19 µm. This high degree of precision is equivalent to that of the reference scanner\*.

Repeat measurements were in the region of 10 µm and the user influence was less than 12 µm. The results were not dependent on the type of preparation.

### 2.1.2 QUADRANT

The images with CEREC Bluecam were taken in auto capture mode and approx. 4–6 exposures were required per quadrant. The software automatically triggers the exposure when the camera is positioned absolutely still above the tooth. CEREC Bluecam demonstrates a significantly improved quadrant precision in comparison to the CEREC 3D camera (34 µm as opposed to 42 µm). The low values of repeat measurements of approx. 13 µm demonstrate the high accuracy of the CEREC Bluecam. The user influence on the precision of the measurement results was extremely low (approx. 15 µm).

### CONCLUSION

The CEREC Bluecam generates digital optical impressions with an unprecedented degree of measurement precision.

## 2.2 MILLING PRECISION

### 2.2.1 CAMERA/MILLING UNIT

The precision of the CEREC system is determined by the resolution of the CEREC camera (25 µm) and the reproducibility of the milling unit ( $\pm 30$  µm). Excluding user-induced influences (e.g. preparation, powdering and exposure technique), the precision of CEREC 3D is in the range  $\pm 55$  µm.

### 2.2.2 MARGINAL FIT OF RESTORATIONS

The marginal accuracy of milled CEREC restorations has continuously improved with each successive software version (from CEREC 1 to current version of CEREC 3D). With regard to the hardware, the introduction of the step bur (tip diameter: 1 mm) represented a major improvement. Within the framework of this multi-centre trial (seven universities) the marginal fit and internal adaptation of CEREC full crowns were measured and compared with laboratory-fabricated ceramic crowns. A group of trained CEREC dentists and a group of non-trained assistants each designed and milled ten molar crowns on the basis of standard models. The crowns (made of the VITA Mark II and Ivoclar ProCad materials) were placed with the aid of Variolink. Empress ceramic crowns sourced from a reputable dental laboratory were also placed.

The marginal fit of the dentist's crowns ( $61.6 \pm 27.9$  µm) and the assistants' crowns ( $60.8 \pm 20.5$  µm) did not differ significantly. The margins of the laboratory-fabricated crowns were slightly wider ( $69.1 \pm 26.9$  µm), which, however, was not statistically significant. With regard to their axial wall adaptation the CEREC crowns were clearly better than the laboratory crowns, whereas in terms of occlusal wall adaptation the laboratory crowns performed better.

OPERATOR	MARGIN	AXIAL WALL
Dentist	$61.8 \pm 27.9$ a	$86.6 \pm 20.9$ b
Assistant	$60.8 \pm 20.5$ a	$88.2 \pm 19.1$ b
Lab Tech	$69.1 \pm 26.9$ a	$125.4 \pm 29.9$ a

Mean values in microns  $\pm$  standard deviation. Groups that are significantly different are indicated by letters P < 0.05.

### CONCLUSION

The marginal fit of CEREC crowns tends to be better than that of laboratory-fabricated ceramic crowns.

\* Laserscan 3D Pro (Willytec, Munich).

Source: Mehl A. Investigation of the optical measurement precision of a new intraoral camera.

Unpublished study carried out by the Department of Computer-aided Restorative Dentistry, Zurich University.

\* This study has been carried out with the CEREC 3 camera. The improved image precision of CEREC Bluecam has therefore not been taken into consideration.

Source: Fasbinder DJ, Multi-Center Trial: Margin Fit and Internal Adaptation of CEREC Crowns. In: Mörmann WH (ed.), State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: Poster

# 3 MARGINAL GAP

## 3.1 ADHESIVE INTERFACE

The performance of the luting materials and the chosen bonding technique have a decisive impact on the success of all-ceramic restorations.

### 3.1.1 MATERIALS

Metal restorations rely principally on macroretention. By contrast etchable all-ceramic materials (silicates/disilicates) are luted directly to the hard dental tissues and rely on microretention. The bonding of CEREC restorations (VITA Mark II, Ivoclar Empress CAD) does not differ from the bonding of laboratory-fabricated inlays, onlays and veneers made of comparable materials. This procedure has remained virtually unchanged since the introduction of dentin adhesives in 1991.

The first step is the CONDITIONING (e.g. etching) of the enamel, dentin and ceramic with the goal of creating a clean micro-roughened surface.

This is followed by the application of a PRIMER, the function of which is to make the clean surface wettable for the hydrophobic bonding material. The third logical step is BONDING – i.e. the application of an unfilled bonding resin, which forms an intermediate layer between the tooth surface, the luting composite and the ceramic material.

Older adhesive systems consist of separate products for each of these steps. The newer systems try to reduce the number of bottles needed.

### CONCLUSION

The adhesive bonding of silicate ceramics has been proved over a period of many years. The various materials must be carefully matched.

High-strength oxide ceramics such as inCeram, aluminium oxide and zirconium oxide do not lend themselves to etching and hence can be conventionally cemented. Self-adhesive luting materials have meanwhile become available.

Luting composites fall into three different categories: chemically cured, light-cured and dual-cured.

This ten-year study compared CEREC 2 inlays which had been luted either with chemically cured or dual-cured composites.

The success rate after ten years was 77% in the case of dual-cured composite and 100% in the case of chemically cured composite.

### CONCLUSION

Dual-cured composites should be used only in situations in which chemically cured or light-cured composites are un-suitable.

### 3.1.2 MARGINAL SEAL

Shortly after the introduction of CEREC there were naturally no long-term studies to draw upon. It was therefore necessary to establish whether the width of the luting interface (i.e. the thickness of the luting composite layer) had any influence on the marginal seal.

All the investigations showed that it was advantageous to locate the restoration margin in the enamel. This *in vitro* study (which involved dye penetration tests) showed that the thickness of the luting composite layer did not have any influence on the marginal seal. Subsequent long-term studies of CEREC 1 and CEREC 2 restorations confirmed these findings.

### CONCLUSION

The thickness of the luting composite layer does not have any influence on the marginal seal.

In relation to deep cavities the question is whether a liner plays a beneficial role for the protection of the pulp.

According to a study carried out by N. Krämer/Erlangen the occurrence of initial hypersensitivity doubled in cases where a liner was laid. The failure rate of ceramic inlays (in this case Empress) trebled when a liner was deployed.

### CONCLUSION

The placement of liners under ceramic inlays/onlays is contra-indicated.

### 3.1.3 WEAR OF THE ADHESIVE INTERFACE

Various Empress inlays placed using Variolink low (low viscosity) and Tetric (high viscosity) were measured in order to determine the wear of the luting composite in highly loaded areas.

After 8 years the mean interfacial width had increased from 176 µm to 207 µm.

The two different composites did not exhibit any significant statistical differences.

### CONCLUSION

Low-viscosity and high-viscosity composites are suitable for the placement of CEREC inlays and onlays.

	Enamel	Dentin	Etchable Ceramic	Non-Etchable Ceramic	Composite
1. Conditioner	35 to 37 % H <sub>3</sub> PO <sub>4</sub>	Self-conditioning Primer	5 % HF	Coe Jet/ Al <sub>2</sub> O <sub>3</sub> powder	
2. Primer	Hydrophobic Bond	Self-conditioning Primer		Organic Silane	
3. Layer-forming Component	Hydrophobic Bond	Pre-cured Amphiphilic Bond		Hydrophobic Bond	
4. Luting Material		Luting Composite			

# 3 MARGINAL GAP

## 3.2 COMPARISON OF OTHER RESTORATION TYPES

Temporaries and methods of bonding labside and chairside produced ceramic inlays have a decisive influence on enamel defects and the margin quality.

### 3.2.1 ENAMEL INTEGRITY


Inlay cavities treated with temporaries demonstrate, after simulated temporary wear in a chewing simulator, a deterioration of the enamel integrity. In particular in both oral and vestibular surfaces cracks developed. By cavities treated immediately with chairside produced ceramic inlays no such enamel defects arose.

### CONCLUSION

Chairside produced CEREC inlays offer the treated tooth a reduced risk of enamel cracks, due to there being no provisional.

### 3.2.2 MARGIN QUALITY

The analysis of bonding systems demonstrated that conventional bonding is still superior to self-adhesive systems. Selective enamel etching as used with CEREC inlays enhances the bond with the hard tooth tissue and improves the quality of the enamel margin. In contrast to the general assumption a broader adhesive gap does not result in inferior margin quality.



Measurement results of damage to the enamel marginal edge (with or without temporaries). Chairside ceramic treated cavities show fewer enamel cracks [TML: thermocycling and mechanical loading].

### CONCLUSION

Immediate treatment without temporaries has a stabilizing effect on the tooth substance. Attainable bonding technology takes into account a broader adhesive gap.

Source: Frankenberger R. Chairside and labside produced ceramic inlays – Influence of temporaries and bonding on the enamel integrity and margin quality. Unpublished study. Publication in preparation.

# 4 OCCLUSAL DESIGN

## 4.1 SOFTWARE

The current version of CEREC 3D includes a variety of tools for mapping the patient's occlusion and articulation (static and functional) and for the automated design of the occlusal surfaces.


- The DENTAL DATABASE contains various sets of data which can be selected according to the specific situation.
- CORRELATION creates a precise and adjustable copy of the existing situation.
- REPLICATION enables the dentist to create an optical impression of any chosen occlusal surface (either contralateral in the patient's mouth or from a separate model). This optical impression can then be placed manually on the preparation.
- ANTAGONIST maps the static occlusion of the antagonists.
- ARTICULATION maps the surface of a functionally generated path (FGP).

By combining DENTAL DATABASE, CORRELATION/REPLICATION with the ANTAGONIST and ARTICULATION tools the dentist is in a position to create functional occlusal surfaces on the computer monitor – manually, semi-automatically or automatically. These occlusal surfaces require only very little subsequent adjustment.

The manual reworking requirement can be reduced from approx. 400 µm (DENTAL DATABASE) to approx. 5 µm (CORRELATION plus ANTAGONIST).

Provided that the equipment parameters are set properly, the dentist can dispense entirely with manual corrections.

To determine the occlusal contact precision of CEREC crowns and laboratory-made Empress crowns respectively, models of natural teeth were measured prior to preparation and after placement of the restorations. This comparison did not reveal any significant differences between the CEREC and Empress crowns.



Number of occlusal contacts of the casts before and after placement of the different crowns (contacts of the restorations excluded!).

### CONCLUSION

Precise occlusal surfaces can be designed on the computer monitor. These require practically no subsequent adjustment in the patient's mouth.

Sources: Fasbinder D J, Predictable CEREC Occlusal Relationships. In Mörmann WH (ed.) State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 93–100 Reich S, Static occlusal precision of two all-ceramic systems. In Mörmann WH (ed.) State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: Poster

# 5 AESTHETICS

## 5.1 POSTERIOR TEETH

CEREC inlays and onlays can be characterized with the help of ceramic stains. After they have been glazed they can be placed in the same way as laboratory-made ceramic restorations.


Due to the special qualities of the CEREC ceramics [chameleon-like shade adaptation; wide choice of lightness, translucency and colour shades], staining is not necessary in most situations.

The CEREC ceramics are easy to polish. In most cases the surface finish is in no way inferior to that of a glazed restoration.

Various studies testify to the good shade adaptation of CEREC ceramics. According to the criteria of the California Dental Association (CDA) 87% of the restorations were rated as excellent. According to the USPHS, the surface characteristics and shade adaptation of all the tested restorations were judged to be excellent or clinically good.

### CONCLUSION

If the ceramic materials are correctly chosen and properly polished, laboratory staining and glazing are unnecessary in most cases.



Mathematical Proportion Guides

Sources: Sjögren G, Molin M, A 10-year prospective evaluation of CAD/CAM-manufactured (CEREC) ceramic inlays cemented with a chemically cured or dual-cured resin composite. Int J Prosthodont; 17/2004: 241–246

## 5.2 ANTERIOR TEETH

After they have been milled CEREC anterior crowns can be stained and glazed. Alternatively, they can be incisally trimmed and then layered using a transparent ceramic material (in cases where especially transparent incisal surfaces are required).

Thanks to their graduated shading intensities, polychromatic blocks (e.g. VITA Triluxe or Empress CAD Multi) make it easier to imitate the natural teeth.

Shading pastes (e.g. VITA Shading Paste, Ivoclar Shade and Stains Kit) and shading powders (e.g. VITA Akzent) permit the rapid characterization of anterior crowns. In simple applications shading and glazing can be combined in a single operation. Multiple firing operations are possible.

The CORRELATION programm allows the shape of the restoration to be simulated prior to milling. In many cases it is possible to create and place chairside anterior crowns during a single appointment. More sophisticated layering techniques are possible.

However, these usually necessitate an indirect procedure using a physical impression and a cast model. Highly complex characterizations can be achieved in this way.

### CONCLUSION

Anterior crowns pose a challenge which can be accomplished during a single appointment with the help of polychromatic blocks and various characterization techniques.

### 5.2.1 VENEERS

More and more CEREC users are offering veneers as part of their treatment repertoire. Partial anterior crowns and veneers are frequently used as a tooth-conserving alternative to a full crown. Characterization can be performed using the methods described for anterior crowns (ceramic stains in combination with transparent ceramic layering materials).

In addition, "background shading" can be used in order to achieve natural-looking results. In this case composite shading materials are applied to the rear surface of the milled veneer. After the veneer has been placed this shading is visible through the thin sliver of ceramic material.

Fine-tuning, contouring and high-gloss polishing are performed after the veneer has been bonded to the tooth. The time input corresponds to that of a CEREC inlay.

### CONCLUSION

CEREC veneers are a fast, tooth-conserving alternative to anterior crowns.

Sources: Masek R, Ultimate CEREC Creations – Comprehensive Single Visit Esthetic Dentistry. In Mörmann WH (ed.) State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 131–138 Reich S, The effect of multicoloured machinable ceramics on the esthetics of all-ceramic crowns. J Prosthet Dent 2002; 88: 44–49 Tsotsos S, Single-appointment, all-ceramic anterior restorations. Int J Comput Dent 2001; 4: 263–272 Wiedhahn K, CEREC Veneers: Esthetics and Longevity. In Mörmann WH (ed.) State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 101–112

# 6 CERAMIC MATERIALS

VITA Mark II is the CEREC material with the longest track record. This feldspar ceramic is available in monochromatic blocks in a variety of 3D Master shades. This same material is also available in a polychromatic version (VITA Triluxe) with differently shaded layers.

The monochromatic CEREC Blocs as well as polychromatic CEREC Blocs PC available from Sirona are also made of feldspar ceramic. They are available in the most popular Classical and 3D Master shades.

The Ivoclar Empress CAD blocks (formerly called ProCad) consist of a leucite-reinforced glass ceramic material. They are available in the shades A-D, with two degrees of translucency respectively. Ivoclar also markets polychromatic blocks ("Multi").

The lithium disilicate glass ceramic blocks (e.max CAD LT) can be conventionally cemented.

## 6.1 STRENGTH

Dental ceramics can be divided into two categories according to their microstructure:

1. Aesthetic enamel-like ceramics with a glass content in excess of 50%. The physical characteristics (e.g. strength, hardness, abrasion properties, opacity, and colour shade) can be modified by the addition of fillers.
  2. Polycrystalline ceramics for frameworks. These consist of particles with an identical crystalline structure. These relatively opaque materials are much stronger than glass ceramics.
- Nearly all these versions are available as conventional laboratory ceramics and as machinable CEREC ceramics. Polycrystalline zirconium oxide and aluminium oxide ceramics are reserved exclusively for CAD/CAM systems.

AESTHETIC CERAMICS	CEREC/inLab
Feldspar	CEREC Blocs, Mark II
Glass/leucite	Empress CAD, Paradigm C
Lithium disilicate	e.max CAD LT, HT
FRAMEWORK CERAMICS	CEREC/inLab
Lithium disilicate	e.max CAD MO
$MgAl_2O_4$ /lanthanum	inCeram Spinell
$Al_2O_3$ /lanthanum	inCeram Alumina
$Al_2O_3/ZrO_3$ /lanthanum	inCeram Zirconia
$Al_2O_3$ (polycrystalline)	inCoris AI, AL-Cubes
$ZrO_2$ Ytt	inCoris ZI, YZ-Cubes, e.max ZirCAD

## CONCLUSION

CEREC and inLab systems can machine all the relevant types of dental ceramics and hence are future-compatible and universally deployable.

# 6 CERAMIC MATERIALS

## 6.2 ABRASION

Veneering ceramics in particular have been criticized for causing wear to the antagonists and for being more abrasion-resistant than tooth enamel.

As long ago as 1991 Krejci demonstrated that VITA Mark II displays the same abrasion properties as natural enamel.

The surfaces of CEREC do not need to be polished or glazed. There is no difference with regard to abrasive wear and tear of the tooth enamel.

### CONCLUSION

CEREC ceramics do not damage the antagonists and display abrasive properties that are similar to gold.

- The abrasiveness of VITA Mark II does not differ significantly from gold. The material itself is abraded at the same rate as gold.
- 1 | ADHESIVE BONDING**
- 1982 MÖRMANN WH: Kompositinlay: Forschungsmodell mit Praxispotential? Quintessenz 1982; 33: 1891–1900.  
MÖRMANN WH, AMEYE C, LUTZ F: Komposit-Inlays: Marginale Adaptation, Randdichtigkeit, Porosität und okklusaler Verschleiss. Dtsch Zahnärztl Z 1982; 37: 438–441.
- 1985 MÖRMANN WH: Keramikinlay – Die Seitenzahnfüllung der Zukunft Vortrag am 30. 3. 1985, Karlsruhe, „25 Jahre Akademie für Zahnärztliche Fortbildung, Karlsruhe“. 4. Internationales Quintessenz-Symposium 1985.  
MÖRMANN WH, BRANDESTINI M, FERRU A, LUTZ F, KREJCI I: Marginale Adaptation von adhäsiiven Porzellainlays in vitro. Schweiz Mschr Zahnmed 1985; 95: 1118–1129.
- 1988 MÖRMANN WH: Innovationen bei ästhetischen Restaurationen im Seitenzahngebiet (Keramik): Computer-gestützte Systeme. Dtsch Zahnärztl Z 1988; 43: 900–903.
- 1989 MÖRMANN WH, BRANDESTINI M, LUTZ F, BARBAKOW F: Chairside computer-aided direkter keramischer Inlays. Quintessence Int 1989; 20: 329–339.  
MÖRMANN WH, BRANDESTINI M: Die CEREC Computer Reconstruction Inlays, Onlays und Veneers. Quintessenz Berlin 1989.
- 1991 BRONWASSER P J, MÖRMANN W H, KREJCI I, LUTZ F: Marginale Adaptation von CEREC-Dicor-MGC-Restaurationen mit Dentinadhäsiven. Schweiz Monatsschr Zahnmed 1991; 101: 162–169.
- FETT H, MÖRMANN W H, KREJCI I, LUTZ F: The effects of short bevels and silanization on marginal adaptation of computer machined mesio-occlusodistal inlays. Quintessence Int 1991; 22: 823–829. MAGNE P, An in Vitro Evaluation of the Marginal and Internal Seals of CEREC Overlays. In Mörmann WH (ed.) International Symposium on Computer Restorations, Berlin: Quintessenz, 1991: 425–440 Schmelz/Dentinhaftung.
- 1993 KÜMIN P, LÜTHY H, MÖRMANN W H: Festigkeit von Keramik und Polymer nach CAD/CIM-Bearbeitung und im Verbund mit Dentin. Schweiz Monatsschr Zahnmed 1993; 103: 1261–1268.  
NOACK MJ, LOCKE LS, ROULET JF: Das Randverhalten adhäsiiv befestigter und mittels Ultraschall eingesetzter Porzellainlays in vivo. Dtsch Zahnärztl Z 1993; 48: 720–723.
- 1994 PEUTZFELDT A: Effect of the ultrasonic insertion technique on the seating of composite inlays. Acta Odontol Scand 1994; 52: 51–54.  
SCHERRER S S, DE RIJK W G, BELSER U C, MEYER J-M: Effect of cement film thickness on the fracture resistance of a machinable glass-ceramic. Dent Mater 1994; 10: 172–177.
- 1995 BESEK M, MÖRMANN W H, PERSI C: Die Aushärtung von Komposit unter zahnfarbenen Inlays. Schweiz Monatsschr Zahnmed 1995; 105: 1123–1128.  
EL-BADRAWY W A, EL-MOWAFY O M: Chemical versus dual curing of resin inlay cements. J Prosthet Dent 1995; 73: 515–524.
- SORENSEN JA, MUNKSGAARD EC: Ceramic inlay movement during polymerisation of resin luting cements. Eur J Oral Sci 1995; 103: 186–189.
- 1996 LOPRESTI JT, DAVID S, CALAMIA JR: Microleakage of CAD-CAM porcelain restorations. Am J Dent 1996; 9: 37–39.  
LUTZ F: State of the art of tooth-colored restoratives. Oper Dent 1996; 21: 237–248.  
MCCOMB D: Adhesive luting cements – classes, criteria, and usage. Comp 1996; 17: 759–774.  
MEHL A, GODESCHA P, KUNZELMANN KH, HICKEL R: Randspaltverhalten von Komposit- und Keramik-inlays bei ausgedehnten Kavitäten. Dtsch Zahnärztl Z 1996; 51: 701–704.
- 1996 MOLIN M K, KARLSSON ST L, KRISTIANSEN M S: Influence of film thickness on joint bend strength of a ceramic/resin composite joint. Dent Mater 1996; 12: 245–249.  
MUNACK J, GEURTSEN W: Marginal adaptation and fit of CEREC I/II and Empress overlays. In: Mörmann WH (ed) CAD/CIM in aesthetic dentistry: CEREC 10 Year Anniversary Symposium. Chicago: Quintessence 1996: 571–579.
- PERDIGAO J, LAMBRECHTS P, VAN MEERBEEK B, BRAEM M, VANHERLE G: Ultramorphological interactions between tooth structure and modern dentin adhesive systems. Academy of Dental Materials, Transactions. 1996; 9: 130–150.
- 1997 GEMALMAZ D, ÖZCAN M, YORUC, ALKUMRU H N: Marginal adaptation of a sintered ceramic inlay system before and after cementation. J Oral Rehabil 1997; 24: 646–651.
- PASHLEY D H, CARVALHO R M: Dentine permeability and dentine adhesion. J Dent 1997; 25: 355–372.
- SCHÜPBACH P, KREJCI I, LUTZ F: Dentin bonding: effect of tubule orientation on hybrid-layer formation. Eur J Oral Sci 1997; 105: 1–9.
- ZELLWEGER U: Der Einfluss der Sauerstoff-Inhibitionsschicht von Dentinadhäsiven auf die Polymerisation verschiedener Abformmaterialien. Zahnmed Diss, Zürich 1997.
- 1998 COX CH F, HAFEZ A A, AKIMOTO N, OTSUKU M, SUZUKI S, TARIM B: Biocompatibility of primer, adhesive and resin composite systems on non-exposed and exposed pulps of non-human primate teeth. Am J Dent 1998; 10: 55–63.
- MEHL A, PFEIFFER A, KREMERS L, HICKEL R: Randständigkeit von CEREC-II-Inlay-Restaurationen bei ausgedehnten Kavitäten mit stark geschwächten Höckern. Dtsch Zahnärztl Z 1998; 53: 57–60.
- MÖRMANN W H, BINDL A, LÜTHY H, RATHKE A: Effects of preparation and luting system on all-ceramic computer-generated crowns. Int J Prosthodont 1998; 11: 333–339.
- PAMEIJER C H, STANLEY H R: The disastrous effects of the „total etch“ technique in vital pulp capping in primates. Am J Dent 1998; 11: 45–54.
- ROSENSTIEL ST F, LAND M F, CRISPIN B J: Dental luting agents: a review of the current literature. J Prosthet Dent 1998; 80: 280–301.
- SHORTALL A C, HARRINGTON E: Temperature rise during polymerization of light-activated resin composites. J Oral Rehabil 1998; 25: 908–913.
- 1999 FRANKENBERGER R, SINDEL J, KRÄMER N, PETSCHELT A: Dentin bond strength and marginal adaptation: direct composite resin vs ceramic inlays. Oper Dent 1999; 24: 147–155.
- MITCHELL CA, DOUGLAS WH, CHENG Y-S: Fracture toughness of conventional, resin-modified glass-ionomer and composite luting cements. Dent Mater 1999a; 15: 7–13.
- MITCHELL CA, PINTADO MR, GEARY L, DOUGLAS WH: Retention of adhesive cement on the tooth surface after crown cementation. J Prosthet Dent 1999b; 81: 668–677.
- PLATT J A: Resin cements: into the 21st century. Comp 1999; 20: 1173–1182.
- 2000 KNOBLOCH LA, KERBY RE, SEGHY R, BERLIN JS, LEE JS: Fracture toughness of resin-based luting cements. J Prosthet Dent 2000; 83: 204–209.
- LANG H: Lichthärtung von Kompositen mit „Plasmalampen“. Dtsch Zahnärztl Z 2000; 55: 154–155.
- 2001 SCHMID C, SCEP S, HEIDEMANN D: Shear strength of composite bonded to porcelain: the influence of a silicone disclosing medium. Int J Comput Dent 2001; 4: 107–116.
- 2002 MUJR IA, FERRARI M: Pulp-dentin biology in restorative dentistry. Part 6: Reactions to restorative materials, tooth-restoration interfaces, and adhesive techniques. Quintessence Int 2002; 33: 35–63.
- 2003 FRANKENBERGER R, KERN M: Dentin Adhesives create a positive bond to dental hard tissue. Int J Comput Dent 2003; 6: 151–162.
- VAN MEERBEEK B, DE MUNCK J, YOSHIDA Y, INOUE S, VARGAS M, VIJAY P, VAN LANDUYT K, LAMBRECHTS P, VANHERLE G: Adhesion to Enamel and Dentin: Current Status and Future Challenges, Oper Dent 2003; 28: 215–235.
- 2005 KOSHIRO K, INOUE S, SANO H, DE MUNK J, VAN MEERBEEK B: In vivo degradation of resin-dentin bonds produced by a self-etch and an etch-and-rinse adhesive. Eur J Oral Sci 2005; 113: 341–348.
- 2006 Krejci I, Bonding of Ceramic Restorations – State of the Art. In Mörmann WH (ed.) State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessenz, 2006: 39–45.

# CEREC PUBLICATIONS

## 2 | FELDSPAR CERAMIC AND GLASS CERAMIC

- 1986 SCHWICKERATH H: Dauerfestigkeit von Keramik. *Dtsch Zahnärztl Z* 1986; 41: 241–266.
- 1990 SEHGI R R, SORENSEN J A, ENGLEMANN M J, ROUMANAS E, TORREES T J: Flexural strength of new ceramic materials. *J Dent Res* 1990; 69: 299, Abstr 1521.
- 1991 KELLY J R, LÜTHY H, GOUGOULAKIS A, POBER R L, MÖRMANN W H: Machining effects on feldspathic and glass ceramic: fractographic analysis. In: Mörmann W H (ed): state of the art of CEREC-method. Chicago: Quintessence 1991: 253–273.
- LUDWIG K: Untersuchungen zur Bruchfestigkeit von Vollkeramikkronen. *Dentallabor* 1991; 5: 647–651.
- 1992 ANUSAVICE KJ: Degradability of dental ceramics. *Adv Dent Res* 1992; 6: 82–89.
- HÖLSCH W, KAPPERT H F: Festigkeitsprüfung von vollkeramischem Einzelzahnersatz für den Front- und Seitenzahnbereich. *Dtsch Zahnärztl Z* 1992; 47: 621–623.
- SCHMID M, FISCHER J, SALK M, STRUB J: Mikrogefüge leucit-verstärkter Glaskeramiken. *Schweiz Monatsschr Zahnmed* 1992; 102: 1046–1053.
- SCHWICKERATH H: Festigkeitsverhalten von CEREC. *Quintessenz* 1992; 43: 669–677.
- 1993 BERNAL G, JONES R M, BROWN D T, GOODACRE C J: The effect of finish line form and luting agent on the breaking strength of Dicor crowns. *Int J Prosthodont* 1993; 6: 286–290.
- KÜMIN P, LÜTHY H, MÖRMANN W H: Festigkeit von Keramik und Polymer nach CAD/CIM-Bearbeitung und im Verbund mit Dentin. *Schweiz Monatsschr Zahnmed* 1993; 103: 1261–1268.
- LÜTHY H, DONG J K, WOHLWEND A, SCHÄRER P: Effects of veneering and glazing on the strength of heat-pressed ceramics. *Schweiz Monatsschr Zahnmed* 1993; 103: 1257–1260.
- 1994 BIENIEK K W, MARX R: Die mechanische Belastbarkeit neuer vollkeramischer Kronen- und Brückenmaterialien. *Schweiz Monatsschr Zahnmed* 1994; 104: 284–289.
- KERN M, FECHTIG T, STRUB J R: Influence of water storage and thermal cycling on the fracture strength of all-porcelain, resin-bonded fixed partial dentures. *J Prosthet Dent* 1994; 71: 251–256.
- 1995 GIORDANO R A, PELLETIER L, CAMPELL S, POBER S: Flexural strength of an infused ceramic, glass ceramic and feldspathic porcelain. *J Prosthet Dent* 1995; 73: 411–418.
- KELLY J R: Perspectives of strength. *Dent Mater* 1995; 11: 103–110.
- MEIER M, FISCHER H, RICHTER E J, MAIER H R, SPIEKERMANN H: Untersuchungen über den Einfluss unterschiedlicher Präparationsgeometrien auf die Bruchfestigkeit vollkeramischer Einzelzahn-Restorationen. *Dtsch Zahnärztl Z* 1995; 4: 295–299.
- 1996 CHAN C, WEBER H: Plaque retention on teeth restored with full-ceramic crowns: a comparative study. *J Prosthet Dent* 1996; 56: 666–671.
- DATZMANN G: CEREC Vitablocs Mark II machinable ceramic. In: Mörmann W H (ed): 10 Year CEREC Anniversary Symposium. Chicago: Quintessence 1996: 205–216.
- DENRY I: Recent advances in ceramics for dentistry. *Crit Rev Oral Biol Med* 1996; 7: 134–143.
- HORNBERGER H, MARQUIS P M, CHRISTIANSEN S, STRUNK H P: Microstructure of a high strength alumina glass composite. *J Mater Res* 1996; 11: 855–858.
- KELLY J R, NISHIMURA J, CAMPBELL S D: Ceramics in dentistry: Historical roots and current perspectives. *J Prosthet Dent* 1996; 75: 18–32.
- LÜTHY H: Strength and Toughness of dental ceramics. In: Mörmann W H (ed): 10 Year CEREC Anniversary Symposium. Chicago: Quintessence 1996: 229–239.
- VAN GOGSWAARDT D C, KOPPITSCH V, MARX R, LAMPERT F: Belastbarkeit von keramischen Einlage-füllungen bei verschiedenen Adhäsivschicht-breiten. *Stomatologie* 1996; 93: 293–296.
- 1997 HICKEL R: Moderne Füllungswerkstoffe. *Dtsch Zahnärztl Z* 1997; 52: 572–585.
- KELLY J R: Ceramics in restorative and prosthetic dentistry. *Annu Rev Mater Sci* 1997; 27: 443–468.
- LAMPE K, LÜTHY H, MÖRMANN W H, LUTZ F: Bruchlast vollkeramischer

Computerkronen. *Acta Med Dent Helv* 1997; 2: 28–35.

ROSENBLUM M A: A review of all-ceramic restorations. *J Am Dent Assoc* 1997; 128: 297–307.

1998 CHEN J-H, MATSUMURA H, ATSUTA M: Effect of different etching periods on the bond strength of a compositeresin to a machinable porcelain. *J Dent* 1998; 26: 53–58.

HÖLAND W, SCHWEIGER M, FRANK M, RHEINBERGER V: A comparison of the microstructure and properties of the IPS Empress 2 and the IPS Empress glass-ceramics. *J Biomed Mater Res* 2000; 53: 297–303.

1999 FISCHER H, MARX R: Mechanische Eigenschaften von Empress 2. *Schweiz Monatschr Zahnmed* 1999; 4: 141–145.

KELLY JR: Clinically relevant approach to failure testing of all-ceramic restorations. *J Prosthet Dent* 1999; 81: 652–661.

2000 TINSCHERT J, ZWEZ D, MARX R, ANUSAVICE KJ: Structural reliability of alumina-, feldspar-, leucite-, mica- and zirconia based ceramics. *J Dent* 2000; 28: 529–535.

2002 REICH S, HORNBERGER H: The effect of multicoloured machinable ceramics on the esthetics of all-ceramic crowns. *J Prosthet Dent* 2002; 88: 44–49.

2004 GUAZZATO M, ALBAKRY M, SWAIN MV: Strength, fracture toughness and microstructure of a selection of all-ceramic materials. Part I. Pressable and alumina glass-infiltrated ceramics. *Dent Mater* 2004a; 20: 441–448.

GUAZZATO M, ALBAKRY M, RINGER SP, SWAIN MV: Strength, fracture toughness and microstructure of a selection of all-ceramic materials. Part II. Zirconia-based dental ceramics. *Dent Mater* 2004b; 20: 441–448.

2006 Kelly R, Machinable Ceramics. In: Mörmann WH (ed.) State of the Art of CAD/CAM-Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 29–38.

## 3 | OCCLUSION, ABRASION, RADIO-OPACITY

- 1991 KREJCI I, LUTZ F, SEBER B, JENSS J: Röntgenopazität von zahnfarbenen Inlaymaterialien und Kompositzementen. *Schweiz Monatsschr Zahnmed* 1991; 101: 299–304.
- 1995 PELKA M, REINELT C, KRÄMER N, FASSBENDER U, PETSCHELT A: In-vivo-Abrasion bei IPS Empress-Inlays. *Dtsch Zahnärztl Z* 1995; 50: 917–919.
- 1996 HAAK R, NOACK M J: Möglichkeiten der Röntgendiagnostik bei Amalgamsatzmaterialien. *Quintessenz* 1996; 47: 1551–1559.
- 1997 AL-HIYASAT A S, SAUNDERS W P, SHARKEY S W, SMITH G MCR, GILMOUR W H: The abrasive effect of glazed, un-glazed, and polished porcelain on the wear of human enamel, and the influence of carbonated soft drinks on the rate of wear. *Int J Prosthodont* 1997; 10: 269–282.
- INOKOSHI S: Posterior restorations: ceramics or composites? *Transactions: Third International Congress on Dental Materials* 1997: 99–108.
- 1998 AL-HIYASAT A S, SAUNDERS W P, SHARKEY S W, SMITH G MCR, GILMOUR W H: Investigation of humanenamel wear against four dental ceramics and gold. *J Dent* 1998; 26: 487–495.
- 1999 FROSCH M: In-vitro-Verschleissuntersuchung moderner Fein-Hybrid-Kompositen gegen Antagonisten aus Schmelz, Gold und Keramik. *Zahnmed Diss Zürich* 1999.
- KREJCI I, ALBERT P, LUTZ F: The influence of antagonist standardization on wear. *J Dent Res* 1999; 78: 713–719.
- MURCHISON D F, CHARLTON D G, MOORE W S: Comparative radio opacity of flowable resin composites. *Quintessence Int* 1999; 30: 179–184.
- 2001 KUNZELMANN KH, JELEN B, MEHL A, HICKEL R: Wear evaluation of M2 100 compared to ceramic CAD/CAM materials. *Int J Comput Dent* 2001; 4: 171–184.
- SCHULTZ C: Passgenauigkeit, marginale Adaptation und okklusale Abrasion von CEREC Inlays. *Med Diss Zürich* 2001.
- 2003 R. POLANSKI, M. LORENZONI, M. HAAS, G. WIMMER, G. ARNETZL: Functional quality of molar crown occlusal surfaces in the different design modes of CEREC 2. *Int J Comput Dent* 2003; 6: 95–101.
- 2004 REICH S: Generation of functional CEREC 3D occlusal surfaces: a comparison of two production methods relevant in practice. *Int J Comput Dent* 2004; 7: 229–238.
- 2005 MEHL A, BLANZ V: New procedure for fully automatic occlusal surface reconstruction by means of a biogeneric tooth model. *J Comput Dent*

2005; 8: 13–25.

2005 MEHL A, BLANZ V, HICKEL R: Biogeneric tooth: a new mathematical representation for tooth morphology in lower first molars. *Eur J Oral Sci* 2005; 113: 333–340.

2006 Fasbinder D J: Predictable CEREC Occlusal Relationships. In: Mörmann WH (ed.) State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 93–100.

KREJCI I, Wear of CEREC and Other Restorative Materials. In: Mörmann WH (ed.) International Symposium on Computer Restorations, Berlin: Quintessence, 1991: 245–251.

REICH S: Static occlusal precision of two all-ceramic systems. In: Mörmann WH (ed.) State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: Poster.

## 4 | CLINICAL AND EXPERIMENTAL FINDINGS

1991 ISENBERG B P, ESSIG M E, LEINFELDER K F, MÜNINGHOFF L A: Clinical evaluation of marginal integrity: two year results. In: Mörmann WH (ed.): International symposium on computer restorations. Chicago: Quintessence 1991: 163–172.

1992 HAAS M, ARNETZL G, WEGSCHEIDER W A, KÖNIG K, BRATSCHKO R: Klinische und werkstoffkundliche Erfahrungen mit Komposit-, Keramik- und Goldinlays. *Dtsch Zahnärztl Z* 1992; 47: 18–22.

HÖGLUND C, VAN DIJKEN J, OLOFSSON A L: A clinical evaluation of adhesively luted ceramic inlays. *Swed Dent J* 1992; 16: 169–171.

MÖRMANN W H, KREJCI I: Computer-designed inlays after 5 years in situ: clinical performance and scanning electron microscopic evaluation. *Quintessence Int* 1992; 23: 109–115.

1994 KREJCI I, FÜLLEMANN J, LUTZ F: Klinische und rasterelektronenmikroskopische Langzeituntersuchung von Kompositinlays. *Schweiz Monatsschr Zahnmed* 1994; 104: 1351–1356.

1995 BRODHECK U, STUDER S, LEHNER C: Sechs Jahre Erfahrung mit einem vollkeramischen Restaurationsystem. *Dentallabor* 1995; 11: 1793–1802.

FLEMMING I, BRONDUM K: A clinical evaluation of porcelain inlays. *J Prosthet Dent* 1995; 74: 140–144.

GLADYS S, VAN MEERBEEK B, INOKOSHI S, WILLEMS G, BRAEM M, LAMBRECHTS P: Clinical and semiquantitative marginal analysis of four tooth-coloured inlay systems at 3 years. *J Dent* 1995; 23: 329–338.

ISIDOR F, BRONDUM K: A clinical evaluation of porcelain inlays. *J Prosthet Dent* 1995; 74: 140–144.

SCHMALZ G, FEDERLIN M, REICH E: Effect of dimension of luting space and luting composite on marginal adaptation of a class II ceramic inlay. *J Prosthet Dent* 1995; 73: 392–399.

1996 BINDL A, MÖRMANN W: Klinische und technische Aspekte der CEREC-InCeram Krone. *Quintessenz* 1996; 47: 775–792.

BRAUNER A W, BIENIEK K W: Seven years of clinical experience with the CEREC inlay system. In: Mörmann W H (ed.): 10 Year CEREC Anniversary Symposium. Chicago: Quintessence 1996: 217–228.

HEYMANN HO, BAYNE SC, STURDEVANT JR, WILDER AD, ROBERSON TM: The clinical performance of CAD-CAM-generated ceramic inlays – a four year study. *J Am Dent Assoc* 1996; 127: 1171–1180.

MÖRIG G: Aesthetic all ceramic restoration: a philosophic and clinical review. *Pract Periodontics Aesthet Dent* 1996; 8: 741–749.

VAN DIJKEN J W V, HÖRSTEDT P: Marginal breakdown of 5-year-old direct composite inlays. *J Dent* 1996; 24: 389–394.

1997 BERG NG, DÈRAND T: A 5-year evaluation of ceramic inlays (CEREC). *Swed Dent J* 1997; 21: 121–127.

BINDL A, MÖRMANN W H: Chairside-Computer-Kronen – Verfahrenszeit und klinische Qualität. *Acta Med Dent Helv* 1997; 2: 293–300.

FRADEANI M, AQUILANO A, BASSEIN L: Longitudinal study of pressed glass-ceramic inlays for four and a half years. *J Prosthet Dent* 1997; 78: 346–353.

FRIEDL K-H, HILLER K-A, SCHMALZ G, BEY B: Clinical and quantitative marginal analysis of feldspathic ceramic inlays at 4 years. *Clin Oral Invest* 1997; 1: 63–168.

HALLER B, POSORSKY A, KLAIBER B: Höckerstabilisierung mit zahnfarbenen Adhäsivinlays in vitro – Einfluß von Inlaymaterial und Total

Bonding. *Dtsch Zahnärztl Z* 1997; 52: 515–519.

ROULET J-F: Longevity of glass ceramic inlays and amalgam – results up to 6 years. *Clin Oral Invest* 1997; 1: 40–46.

1997 THONEMANN B, FEDERLIN M, SCHMALZ G, SCHAMS A: Clinical evaluation of heat-pressed glass-ceramic inlays in vivo: 2-year results. *Clin Oral Invest* 1997; 1: 27–34.

1998 FELDEN A, SCHMALZ G, FEDERLIN M, HILLER K-A: Retrospective clinical investigation and survival analysis on ceramic inlays and partial ceramic crowns: results up to 7 years. *Clin Oral Invest* 1998; 2: 161–167.

HAYASHI M, TSUCHITANI Y, MIURA M, TAKESHIGE F, EBISU S: 6-year clinical evaluation of fired ceramic inlays. *Oper Dent* 1998; 23: 318–326.

HOFERE: In-vivo-Langzeitverhalten von zahnfarbenen Restaurationen im Seitenzahnbereich. *Zahnmed Diss*, Zürich 1998.

JACKSON R D: Asthetische Inlays und Onlays: eine restaurative Alternative. *Phillip J* 1998; 15: 169–173.

KERSCHAUM T: Langzeit-Erfolgsaussichten von festsetzendem Zahnersatz. *Zahnärztl Mitt* 1998; 88: 64–69.

KREJCI I, LUTZ F, MÖRMANN WH: Zahnfarbene Adhäsive Restaurationen im Seitenzahnbereich. *Eigenverlag PPK*, Zürich 1998.

LEHNER C, STUDER S, BRODHECK U, SCHÄRER P: Six-year clinical results of leucite-reinforced glass ceramic inlays and onlays. *Acta Med Dent Helv* 1998; 3: 137–146.

MÖRMANN WH, BINDL A: Die CEREC-Computerkrone – erste klinische und wissenschaftliche Erfahrungen. *Dental Magazin* 1998; 16: 82–91.

SCHEIBENBOGEN A, MANHART J, KUNZELMANN K-H, HICKEL R: One-year clinical evaluation of composite and ceramic inlays in posterior teeth. *J Prosthet Dent* 1998; 80: 410–416.

SJÖGREN G, MOLIN M, VAN DIJKEN JWV: A 5-year clinical evaluation of ceramic inlays (CEREC) cemented with a dual-cured or chemically cured resin composite luting agent. *Acta Odontol Scand* 1998; 56: 263–267.

1999 BINDL A, MÖRMANN W H: Clinical evaluation of adhesively placed CEREC Endo-Crowns after 2 years – Preliminary results. *J Adhesive Dent* 1999; 1: 255–265.

DONLY K J, JENSEN M E, TRIOLI P, CHAN D: A clinical comparison of resin composite inlay and onlay posterior restorations and cast gold restorations at 7 years. <

# CEREC PUBLICATIONS

- computer-aided machined restorations. *Dent Clin N Am* 2002; 46: 405–426.
- 2003 BINDL A, MÖRMANN WH: Clinical and SEM evaluation of all-ceramic chairside CAD/CAM-generated partial crowns. *Eur J Oral Sci* 2003; 111: 163–169.
- POSSELT A, KERSCHBAUM T: Longevity of 2328 chairside CEREC inlays and onlays. *Int J Comput Dent* 2003; 6: 231–248.
- 2004 Sjögren G, Molin M, A 10-year prospective evaluation of CAD/CAM – manufactured (CEREC) ceramic inlays cemented with a chemically cured or dual-cured resin composite. *Int J Prosthodont* 2004; 17: 241–246.
- 2005 Bindl A, Survival of Ceramic Computer-aided Design/Manufacturing Crowns Bonded to Preparations with Reduced Macorentention Geometry. *Int J Prsthodont*; 18, 2005: 219–224.
- 2006 Arnetzl G, Different Ceramic Technologies in a Clinical Longterm Comparison. In Mörmann WH [ed.] State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 65–72.
- Kerschbaum T, A Comparison of the Longevity and Cost-effectiveness of Three Inlay-types. In Mörmann WH [ed.] State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 73–82.
- Otto T, Computer-Aided Direct All-Ceramic Crowns: 4 Year Results. In Mörmann WH [ed.] State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: Poster.
- Reiss B, Eight-Year Clinical Study in a Dental Practice. In Mörmann WH [ed.] State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 57–64.
- Wiedhahn K, CEREC Veneers: Esthetics and Longevity. In Mörmann WH [ed.] State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 101–112.
- 2008 Otto T, Schneider D: Long-Term Clinical Results of Chairside CEREC CAD/CAM Inlays and Onlays: A Case Series. *Int J Prosthodont* 2008; 21: 53–59.
- 2009 Frankenberger R, Chairside und labside gefertigte Keramikinlays – Einfluss von Provisorien und Adhäsivtechnik auf die Schmelzintegrität und Randqualität. Unveröffentl. Studie. Publikation in Vorbereitung.
- 5 | RESTORATIVE TECHNOLOGY**
- 1978 LUTZ F, LEUTHARD P: Verschleissfeste MOD-Kompositfüllungen durch Einpolymerisation von zentrischen Stops aus Keramik – 4-Jahres-Resultate. *Schweiz Monatsschr Zahnheilk* 1978; 88: 739–752.
- 1982 MÖRMANN W H: Kompositinlay: Forschungsmodell mit Praxispotential? *Quintessenz* 1982; 33: 1891–1900.
- 1985 LUTZ F, MÖRMANN W, KREJCI I: Seitenzahnkomposite – Ja, Nein oder Jein? *Dtsch Zahnärztl Z* 1985; 40: 892–896.
- 1987 JENSEN ME, REDFORD DA, WILLIAMS BT, GARDNER F: Posterior Etched-Porcelain Restorations: An in Vitro Study. *Compend Contin Educ Dent* 1987; 7: 615–622.
- LUTZ F, KREJCI I, MÖRMANN W: Die zahnfarbene Seitenzahnrestoration. *Phillip J* 1987; 3: 127–137.
- MÖRMANN W H, BRANDESTINI M, LUTZ F: Das CEREC-System: Computer-gestützte Herstellung direkter Keramikinlays in einer Sitzung. *Quintessenz* 1987; 38: 457–470.
- 1988 DURET F, BLOUIN JL, DURET M: CAD/CAM in dentistry. *J Am Dent Assoc* 1988; 117: 715–720.
- MÖRMANN W H: Innovationen bei ästhetischen Restaurierungen im Seitenzahngebiet (Keramik): Computer-unterstützte Systeme. *Dtsch Zahnärztl Z* 1988; 43: 900–903.
- 1989 MÖRMANN W H, BRANDESTINI M: Die CEREC Computer Rekonstruktion – Inlays, Onlays und Veneers. *Quintessenz*, Berlin 1989.
- 1990 HÜRZELER M, ZIMMERMANN E, MÖRMANN WH: Marginale Adaptation von maschinell hergestellten Onlays in vitro. *Schweiz Monatsschr Zahnmed* 1990; 100: 715–719.
- JÄGER K, HENZ B, WIRZ J, GRABER G: Marginale Passgenauigkeit befestigter adhäsiver Keramikinlays. *Schweiz Monatsschr Zahnmed* 1990; 100: 1304–1309.
- KAPPERT H F, KNODE H: InCeram auf dem Prüfstand. *Quintessenz Zahntech* 1990; 8: 980–1002.
- LEVY H, DANIEL X: Working with the InCeram porcelain system. *Prothèse Dentaire* 1990; 44–45: 1–11.
- WOHLWEND A, SCHÄRER P: Die Empress-Technik, ein neues Verfahren zur Herstellung von vollkeramischen Kronen, Inlays und Facetten. *Quintessenz Zahntech* 1990; 8: 966–978.
- 1991 MÖRMANN WH [ed.]: Proceedings of the international symposium on computer restorations. Chicago: Quintessence 1991: 23–548.
- REKOW E D: Dental CAD-CAM systems: What is the state of the art? *J Am Dent Assoc* 1991; 122: 43–48.
- ROULET J F, NOACK M J: Tooth coloured conventional and CEREC restoration – claim and reality. In: Mörmann WH [ed.]: International symposium on computer restorations. Chicago: Quintessence 1991: 233–243.
- 1992 DONG J K, LÜTHY H, WOHLWEND A, SCHÄRER P: Heat-pressed ceramics: Technology and strength. *Int J Prosthodont* 1992; 5: 9–16.
- FUTTERKNECHT N, JINOIAN V: A renaissance of ceramic prosthetics? *Quintessenz Dent Tech* 1992; 15: 65–78.
- PRÖBSTER L, DIEHL J: Slip-casting alumina ceramics for crown and bridge restorations. *Quintessenz Int* 1992; 23: 25–31.
- SCHMID M, FISCHER J, SALK M, STRUB J: Mikrogefüge leucit-verstärkter Glaskeramiken. *Schweiz Monatsschr Zahnmed* 1992; 102: 1046–1053.
- SIEBER C: Variationen der Lichtleitfähigkeit und Leuchtkraft. *Quintessenz Zahntech* 1992; 18: 1123–1133.
- FURRER O, MÖRMANN W H: Effizienz und Kantenqualität beim computertechnischen Formschleifen von Keramikinlays. *Schweiz Monatsschr Zahnmed* 1993; 103: 851–859.
- GREY N F, PIDDOCK V, WILSON M A: In vitro comparison of conventional crowns and a new all ceramic system. *J Dent* 1993; 21: 47–51.
- 1994 EIDENBENZ S, LEHNER CH R, SCHÄRER P: Copy milling ceramic inlays from resin analogs: a practical approach with the Celay system. *Int J Prosthodont* 1994; 7: 134–142.
- NATHANSON D, RIIS D N, CATALDO G L, ASHAYERI N: CAD/CAM ceramic inlays and onlays: using an indirect technique. *J Am Dent Assoc* 1994; 125: 421–427.
- SIEBER C, THIEL N: Eine lichtoptische Möglichkeit – Spinell Luminaries. *Quintessenz Zahntech* 1994; 20: 1041–1051.
- STRUPOWSKY M: Das Precedent-DCS-System, numerisch gesteuerte Zahntechnik. *Dental Labor* 1994; 42: 1809–1815.
- 1995 HEGENBARTH E A: Die Symbiose aus Computertechnologie und Kreativität. *Dental Labor* 1995; 43: 797–809.
- NASEDKIN J N: Ceramic inlays and onlays: update 1995. *J Can Dent Assoc* 1995; 61: 676–682.
- SEGH R R, SORENSEN J A: Relative flexual strength of six new ceramic materials. *Int J Prosthodont* 1995; 8: 239–246.
- 1996 FUTTERKNECHT N: Renaissance in der Vollkeramik? [I]. *Quintessenz Zahntech* 1996; 16: 1185–1197.
- LÜTHY H, PIETROBON N, SISERA M, WOHLWEND A, LOEFFEL O: White esthetics. *Schweiz Monatsschr Zahnmed* 1996; 106: 897–904.
- MÖRMANN W H [ed.]: 10 Year CEREC Anniversary Symposium. Chicago: Quintessence 1996: 21–663.
- MÖRMANN W H, BRANDESTINI M: The fundamental inventive principles of CEREC CAD/CIM and other CAD/CAM methods. In: Mörmann W H [ed.]: 10 Year CEREC Anniversary Symposium. Chicago: Quintessence 1996; 81–110.
- PFEIFFER J: The Character of CEREC 2. In: Mörmann WH [Hrsg.]: CAD/CIM in aesthetic dentistry: CEREC 10 Year Anniversary Symposium. Chicago: Quintessence 1996: 255–265.
- SIEBERT G K, NEUKIRCHEN S: Full-ceramic restorations – A survey of different systems. In: Mörmann W H [ed.]: 10 Year CEREC Anniversary Symposium. Chicago: Quintessence 1996: 73–80.
- TOUATI B: The evolution of aesthetic restorative materials for inlays and onlays: a review. *Int Aesth Chron* 1996; 8: 657–666.
- VAN PELET A W J, DE KLOET H J, VAN DER KUY P: Keramische inlays en onlays. *Ned Tijdschr Tandheelkd* 1996; 103: 472–476.
- 1997 FELBER, L., LEEBACH, TH., MÖRMANN, W. H.: Computergestützte vollautomatische Konstruktion von Inlays. Eine qualitative Analyse. *Acta Med Dent Helv* 1997; 2: 217–225.
- HICKEL R., DASCH W, MEHL A, KREMERS L: CAD/CAM – fillings of the future? *Int Dent J* 1997; 47: 247–258.
- MÖRMANN W H, BINDL A: The new creativity in ceramic restorations: Dental CAD/CIM. *Quintessence Int* 1997; 27: 821–828.
- PRÖBSTER L: Zum heutigen Stand vollkeramischer Restaurierungen. *Zahnärztl Mitt* 1997; 20: 44–50.
- SCHLUDERER M: CEREC im Praxislabor. *Dental Magazin* 1997; 3: 42–44.
- 1998 CHALIFOUX P R: Treatment considerations for posterior laboratory-fabricated composite resin restorations. *Pract Periodont Aesthet Dent* 1998; 10: 969–978.
- KREJCI I, DIETSCHI D, LUTZ F U: Principles of proximal cavity preparation and finishing with ultrasonic diamond tips. *Pract Periodont Aesthet Dent* 1998; 10: 295–298.
- KREJCI I, LUTZ F, MÖRMANN WH: Zahnfarbene adhäsive Restaurierungen im Seitenzahnbereich. Eigenverlag PPK, Zürich 1998.
- LANG H, SCHÜLER N, NOLDEN R: Keramikinlay oder Teilkrone? *Dtsch Zahnärztl Z* 1998; 53: 53–56.
- 1998 MC KAREB E A: All-Ceramic alternatives to conventional metal-ceramic restorations. *Comp* 1998; 19: 307–325.
- SCHMIDT A, WALTER M, BÖNING K: CAD/CAM/CIM-Systeme in der restaurativen Zahnmedizin. *Quintessenz* 1998; 49: 1111–1122.
- 1999 BINDL A, WINDISCH S, MÖRMANN W H: Full-Ceramic CAD/CIM Anterior Crowns and Copings. *Int J Comp Dent* 1999; 2: 97–111.
- ESTEFAN D, DAVID A, DAVID S, CALAMIA J: A new approach to restorative dentistry: fabricating ceramic restorations using CEREC CAD/CAM. *Compend Contin Educ Dent* 1999; 20: 555–560.
- HUGO B: Oszillierende Verfahren in der Präparationstechnik [Teil I]. *Schweiz Monatsschr Zahnmed* 1999; 109: 140–153.
- MAGNE P, DOUGLAS W H: Additive contour of porcelain veneers: a key element in enamel preservation, adhesion, and esthetics for aging dentition. *J Adhesive Dent* 1999; 1: 81–92.
- MAGNE P, DOUGLAS W H: Rationalization of esthetic restorative dentistry based on biomimetics. *J Esthet Dent* 1999; 11: 5–15.
- MAGNE P, DOUGLAS W H: Porcelain veneers: dentin bonding optimization and biometric recovery of the crown. *Int J Prosthodont* 1999; 12: 111–121.
- MEHL A, HICKEL R: Current state of development and perspectives of machine-based production methods for dental restorations. *Int J Comput Dent* 1999; 2: 9–35.
- MÖRMANN W H, BINDL A, RICHTER B, APHOLT W, TOTH RT: CEREC computer aided design, computer integrated manufacturing: full ceramic crowns. In: Mörmann W H [ed.]: CAD-CIM Library, SFCZ Publishing, Zurich 1999.
- PFEIFFER J: Dental CAD/CAM Technologies: The optical impression [II]. *Int J Comput Dent* 1999; 2: 65–72.
- POLANSKY R, ARNETZL G, SMETAN M, HAAS M, LORENZONI M: The Production of CEREC restorations from a plaster cast. *Int J Comput Dent* 1999; 2: 37–44.
- VAN DER ZEL J M: Heutige CAD/CAM-Systeme im Vergleich. *Quintessenz Zahntech* 1999; 2: 193–204.
- ZITZMANN NU, MARINELLO CP, LÜTHY H: The Procera Allceram all-ceramic system. The clinical and technical laboratory aspects in the use of a new all-ceramic system. *Schweiz Monatsschr Zahnmed* 1999; 109: 820–834.
- 2000 MAGNE P, DOUGLAS W H: Cumulative effects of successive restorative procedures on anterior crown flexure: intact versus veneered incisors. *Quintessenz Int* 2000; 31: 5–18.
- MÖRMANN W H, BINDL A: CEREC 3 – ein Quantensprung bei Computer-Restaurierungen. Erste klinische Erfahrungen. *Quintessenz* 2000; 51: 157–171.
- 2001 APHOLT W, BINDL A, LÜTHY H, MÖRMANN W: Flexural strength of CEREC 2 machined and joined together InCeram-Alumina and InCeram-Zirconia bars. *Dent Mater* 2001; 17: 260–267.
- CHRISTENSEN G J: Computerized restorative dentistry. State of the art. *J Am Dent Assoc* 2001; 132: 1301–1303.
- FILSER F, KOCHER P, WEIBEL F, LÜTHY H, SCHÄRER P, GAUCKLER LJ: Reliability and strength of all-ceramic dental restorations fabricated by direct ceramic machining (DCM). *Int J Comput Dent* 2001; 4: 89–106.
- HEHN S: The evolution of a chairside CAD/CAM system for dental restorations. *Compend Contin Educ Dent* 2001; 22 [6 Suppl]: 4–6.
- JEDYNAKIEWICZ NM, MARTIN N: Functionally generated pathway theory, application and development in CEREC restorations. *Int J Comput Dent* 2001; 4: 5–36.
- KURBAD A: CEREC goes inLab – the metamorphosis of the system. *Int J Comput Dent* 2001; 4: 125–143.
- PREVOST AP, BOUCHARD Y: CEREC: correlation, an accurate and practical method for occlusal reconstruction. *Jnt J Comp Dent* 2001; 4: 185–193.
- SUTTOR D, BUNKE K, HOESCHELER S, HAUPTMANN H, HERTLEIN G: LAVA – the system for all-ceramic ZrO<sub>2</sub> crown and bridge frameworks. *Int J Comput Dent* 2001; 4: 195–206.
- WIEDHAHN K: Function with registration: Simple, fast, and safe using a new registration material. *Int J Comp Dent* 2001; 4: 207–216.
- 2002 BISLER A, BOCKHOLT U, KORDASS B, SUCHAN M, VOSS G: Der virtuelle Artikulator. *Int J Comput Dent* 2002; 5: 101–106.
- CASTANO MC, ZAPATA U, PEDROZA A, JARAMILLO JD, ROLDAN S: Creation of a three-dimensional model of the mandible and the TMJ in vivo by means of the finite element method. *Int J Comput Dent* 2002; 5: 87–99.
- EDELHOFF D, SORENSEN JA: Tooth structure removal associated with various preparation designs for posterior teeth. *Int J Periodont Rest* 2002; 22: 241–249.
- 2002 HIKITA K, UCHIYAMA Y, OTSUKI M, IYAMA K, DURET F: Function and clinical application of dental CAD/CAM "GN-1". *Int J Comput Dent* 2002; 5: 11–23.
- 2003 GÄRTNER C, KORDASS B: The virtual articulator: development and evaluation. *Int J Comput Dent* 2003; 6: 11–23.
- 2004 GLAUSER R, SAILER I, WOHLWEND A, STUDER S, SCHIBLI M, SCHÄRER P: Experimental zirconia abutments for implant-supported single-tooth restorations in esthetically demanding regions: 4-year results of a prospective clinical study. *Int J Prosthodont* 2004; 17: 285–290.
- MÖRMANN WH: The origin of the CEREC method: a personal review of the first 5 years. *Int J Comput Dent* 2004; 7: 11–24.
- TINSCHERT J, NATT G, HASSENPLUG S, SPIEKERMANN H: Status of current CAD/CAM technology in dental medicine. *Int J Comput Dent* 2004; 7: 25–45.
- 2009 Mehl A: Genauigkeitsuntersuchung zur optischen Vermessung mit einer neuen intraoralen Kamera. Unveröffentl. Studie an der Station für computergestützte restaurative Zahnmedizin, Universität Zürich.
- 6 | MARGINAL FIT**
- 1985 DEMDON HW: Ability to evaluate nonvisible margin with an explorer. *Oper Dent* 1985; 10: 6–11.
- 1990 JÄGER K, HENZ B, WIRZ J, GRABER G: Marginale Passgenauigkeit befestigter adhäsiver Keramikinlays. *Schweiz Monatsschr Zahnmed* 1990; 100: 1304–1309.
- 1992 INOKOSHI S, VAN MEERBEK B, WILLEMS G, LAMBRECHTS P, BREM M, VANHERLE: Marginal accuracy of CAD/CAM inlays made with the original and the updated software. *J Dent* 1992; 20: 171–177.
- STEUBER-SANDERA I: Fügeverhältnisse bei computergestützter Herstellung von Keramikinlays unter Anwendung von drei Gerätenerationen. *Med Diss*, Zürich 1992.
- 1993 KREJCI I, LUTZ F, REIMER M: Marginal adaptation and fit of adhesive ceramic inlays. *J Dent* 1993; 21: 39–46.
- MOLIN M, KARLSSON S: The fit of gold inlays and three ceramic inlay systems – A clinical and in vitro study. *Acta Odontol Scand* 1993; 51: 201–206.
- 1994 KAWAI K, ISENBERG B P, LEINFELDER K F: Effect of gap dimension on composite resin cement wear. *J Dent Res* 1994; 25: 53–58.
- SIERVO S, PAMPALONE A, SIERVO P, SIERVO R: Where is the gap? Machinable ceramic systems and conventional laboratory restorations at a glance. *Quintessenz Int* 1994; 25: 773–779.
- 1995 RINKE S, HÜLS A, JAHN L: Marginal accuracy and fracture strength of conventional and copy-milled all-ceramic crowns. *Int J Prosthodont* 1995; 8: 303–310.
- SCHMALZ G, FEDERLIN M, REICH E: Effect of dimension of luting space and luting composite on marginal adaptation of a class II ceramic inlay. *J Prosthet Dent* 1995; 73: 392–399.
- SHINKAI K, SUZUKI S, LEINFELDER K F, KATOH Y: Effect of gap dimension on wear resistance of luting agents. *Am J Dent* 1995; 8: 149–151.

# CEREC PUBLICATIONS

- 8 | CEREC CROWNS**
- 1996 SJÖGREN G: Marginal and internal fit of four different types of ceramic inlays after luting. An *in vitro* study. *Acta Odontol Scand* 1995; 53: 24–28.
- SORENSEN J A, MUNKSGAARD E CH: Interfacial gaps of resin cemented ceramic inlays. *Eur J Oral Sci* 1995; 103: 116–120.
- 1996 MUNACK J, GEURTSSEN W: Marginal adaptation and fit of CEREC I/II and Empress overlays. In: Mörmann WH (ed) CAD/CIM in aesthetic dentistry: CEREC 10 Year Anniversary Symposium. Chicago: Quintessence 1996: 571–579.
- VAN GOESWAARDT D C, KOPPITSCH V, MARX R, LAMPERT F: Belastbarkeit von keramischen Einlage-füllungen bei verschiedenen Adhäsivschicht-breiten. *Stomatologie* 1996; 93: 293–296.
- 1997 GUZMAN A F, MOORE B K, ANDRES C J: Wear resistance of four luting agents as a function of marginal gap distance, cement type, and restorative material. *Int J Prosthodont* 1997; 10: 415–425.
- MÖRMANN W H, SCHUG J: Grinding precision and accuracy of fit of CEREC 2 CAD-CIM inlays. *J Amer Dent Assoc* 1997; 128: 47–53.
- SJÖGREN G, HEDLUND S-O: Filler content and gap width after luting of ceramic inlays, using the ultrasonic insertion technique and composite resin cements. *Acta Odontol Scand* 1997; 55: 403–407.
- SULAIMAN F, CHAI J, JAMESON L M, WOZNIAK W T: A comparison of the marginal fit of InCeram, IPS Empress and Procera crowns. *Int J Prosthodont* 1997; 10: 478–484.
- UNTERBRINK G L: Differenzierende Analyse der erforderlichen Genauigkeit bei laborgefertigten Restaurationen. *Phillip J* 1997; 11–12: 386–388.
- 1998 FISCHER J, GRÄBER H-G, LAMPERT F: Die Randschlussqualität von zahnfarbenen Klasse-II-Versorgungen. *ZWR* 1998; 107: 192–195.
- MAY KB, RUSSEL MM, RAZZOOG ME, LANG BR: Precision of fit: the Procera AllCeram crown. *J Prosthet Dent* 1998; 80: 394–404.
- 2001 SCHULTZ CH: Passgenauigkeit, marginale Adaptation und okklusale Abrasion von CEREC Inlays. *Med Diss*, Zürich: 2001.
- 2004 LUTHARDT RG, BORNEMANN G, LEMELSON S, WALTER MH, HULS A: An innovative method for evaluation of the 3-D internal fit of CAD/CAM crowns fabricated after direct optical versus indirect laser scan digitizing. *Int J Prosthodont* 2004; 17: 680–685.
- 2006 Fasbinder DJ, Multi-Center Trial: Margin Fit and Internal Adaptation of CEREC Crowns. In: Mörmann WH (ed.), State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: Poster.
- 2008 Reich S, Gozdowski S, Trentzsch L, Frankenberger R, Lohbauer U: Marginal Fit of Heat-pressed vs CAD/CAM Processed All-ceramic Onlays Using a Milling Unit Prototype. *Operative Dentistry* 2008; 33-6: 644–650.
- 7 | FINISHING, POLISHING, REPAIRS**
- 1991 SCHMID O, KREJCI I, LUTZ F: Ausarbeitung von adhäsiven zahnfarbenen Inlays aus Komposit und Keramik. *Schweiz Monatsschr Zahnmed* 1991; 101: 177–184.
- 1995 FEHER A, MÖRMANN W: Die Ausarbeitung von Keramik-Restaurationen mit superfeinen Diamantinstrumenten. *Schweiz Monatsschr Zahnmed* 1995; 105: 474–479.
- KREJCI I, LUTZ F, KREJCI D: Schall- / ultraschallbetriebene diamantierte Instrumente zur Kavitätenpräparation zum Konturieren und zum Finieren. *ZWR* 1995; 104: 781–786.
- MALONE W F P, TAYLOR TH D: Effects of ultrasonic scaling and periodontal curettage on surface roughness of porcelain. *J Prosthet Dent* 1995; 73: 227–232.
- 1996 KREJCI I, LUTZ F, BAFELLI G, KILCHER B: Nondestructive Politur mit einem neu entwickelten rotierenden Bürstchen. *ZWR* 1996; 105: 304–308.
- 1997 BOLLEN CM, LAMBRECHTS P, QUIRYNEN M: Comparison of surface roughness of oral hard materials to the threshold surface roughness for bacterial plaque retention: a review of the literature. *Dent Mater* 1997; 13: 258–69.
- 1998 DENEHY G, BOUSCHLICHER M, VARGAS M: Intraoral repair of cosmetic restorations. *Dent Clin North Am* 1998; 42: 719–737.
- 2001 TSOTSOS S: An indirect/direct method for contouring and finishing CEREC restorations. *Int J Comput Dent* 2001; 4: 37–45.

- 8 | CEREC 3D APPLICATIONS AND CLINICAL ISSUES**
- 1996 BINDL A, MÖRMANN WH: Clinical and technical aspects of the CEREC InCeram crown. In: Mörmann, W. H. (ed): CAD/CIM in Aesthetic Dentistry. CEREC 10 Year Anniversary Symposium. Chicago: Quintessence 1996: 441–462.
- BINDL A, MÖRMANN WH: Klinische und technische Aspekte der CEREC InCeram Krone. *Quintessenz* 1996; 47: 775–792.
- BITZ M, MÖRMANNWH: Documentation of CEREC crowns via electronic communication. In: Mörmann, W. H. (ed): CAD/CIM in Aesthetic Dentistry. CEREC 10 Year Anniversary Symposium. Chicago: Quintessence 1996: 495–516.
- LAMPE K, LÜTHY H, MÖRMANN WH: Fracture load of all-ceramic computer crowns. In: Mörmann, W. H. (ed): CAD/CIM in Aesthetic Dentistry. CEREC 10 Year Anniversary Symposium. Chicago: Quintessence 1996: 463–482.
- 1997 BINDL A, MÖRMANN WH: Chairside Computer-Krone – Verfahrenszeit und klinische Qualität. *Acta Med Dent Helv* 2: 1997; 2: 293–300.
- DEVIGUS A: Die CEREC-2-Frontzahnkrone. *Dental Magazin* 1997; 3: 38–41.
- LAMPE K, LÜTHY H, MÖRMANN WH, LUTZ F: Bruchlast vollkeramischer Computerkronen. *Acta Med Dent Helv* 1997; 2: 76–83.
- 1998 MÖRMANN WH, RATHKE A, LÜTHY H: Der Einfluss von Präparation und Befestigungsmethode auf die Bruchlast vollkeramischer Computerkronen. *Acta Med Dent Helv* 1998; 3: 29–35.
- MÖRMANN WH, BINDL A: Die CEREC-Computerkrone – erste klinische und wissenschaftliche Erfahrungen. *Dental Magazin* 1998; 16: 82–91.
- MÖRMANN WH, BINDL A, LÜTHY H, RATHKE A: Effects of preparation and luting system on all-ceramic computer-generated crowns. *Int J Prosthodont* 1998; 11: 333–339.
- 1999 BINDL A, WINDISCH S, MÖRMANN WH: Full-Ceramic CAD/CIM Anterior Crowns and Copings. *Int J Comput Dent* 1999; 2: 97–111.
- 1999 BINDL A, MÖRMANN WH: Clinical evaluation of adhesively placed CEREC Endo-Crowns after 2 years – Preliminary results. *J Adhesive Dent* 1999; 1, 255–265.
- CHEN HY, HICKEL R, SETCOS JC, KUNZLEMMAN KH: Effects of surface finish and fatigue testing on the fracture strength of CAD/CAM and pressed ceramic crowns. *J Prostet Dent* 1999; 82: 468–475.
- MALAMENT KA, SOCRANSKY SS: Survival of Dicor glass-ceramic dental restorations over 14 years: Part I. Survival of Dicor complete coverage restorations and effect of internal surface acid etching, tooth position, gender and age. *J Prosthet Dent* 1999; 81: 23–32.
- MÖRMANN WH, BINDL A, RICHTER B, APHOLT W, TOTH RT: CEREC Computer Aided Design, Computer Integrated Manufacturing: full-ceramic crowns. CAD/CIM Library, Vol 2, Zurich: SFCZ Publishing 1999.
- TOTH RT, MÖRMANN W: CEREC 2 vollkeramische CAD/CIM Computerkronen – Konstruktion im Front- und Seitenzahnbereich. CAD/CIM Bibliothek Band 1, SFCZ-Verlag Zürich 1999.
- WINDISCH S, BINDL A, MÖRMANN WH: Passgenauigkeit von vollkeramischen CEREC CAD/CIM Frontzahnkronen und Frontzahnkronenkappen. *Acta Med Dent Helv* 1999; 3: 29–37.
- 2000 CHAI J, TAKAHASHI Y, SULAIMAN F, CHONG K, LAUTENSCHLAGER EP: Probability of fracture of all-ceramic crowns. *Int J Prosthodont* 2000; 13: 420–424.
- MÖRMANN WH, BINDL A: The CEREC 3 – A quantum leap for computer-aided restorations: Initial clinical results. *Quintessence Int* 2000; 31: 699–712.
- MÖRMANN WH, BINDL A, APHOLT W: CEREC 3 Computer-Restaurationen – erste klinische Erfahrungen. *Zahnärztliche Mitteilungen* 2000; 90: 2860–2869.
- POLANSKY R, ARNETZL G, HAAS M, KEIL C, WIMMER G, LORENZONI M: Restdentinstärke nach 1,2 mm-Stufenpräparation für CEREC-Kronen. *Int J Comput Dent* 2000; 3: 243–258.
- 2001 BESIMO CE, SPIELMANN H-P, ROHNER H-P: Computer-assisted generation of all-ceramic crowns and fixed partial dentures. *Int J Comput Dent* 2001; 4: 243–262.
- TSOTSOS S: Single-appointment, all-ceramic anterior restorations. *Int J Comput Dent* 2001; 4: 263–272.
- 9 | CEREC 3D APPLICATIONS AND CLINICAL ISSUES**
- 2000 HANSEN S: Präparationen für CEREC 3: Gibt es noch Einschränkungen? *Int J Comput Dent* 2000; 3: 197–205.
- 2001 TSOTSOS S: Single-appointment, all-ceramic anterior restorations. *Int J Comput Dent* 2001; 4: 263–272.
- 2003 ENDER A, WIEDHAHN K, MÖRMANN WH: Chairside multi-unit restoration of a quadrant using the new CEREC 3D-Software. *Int J Comput Dent* 2003; 6: 89–94.
- FRITSCH G: Correlation and Function in CEREC 3D – what is new? *Int J Comput Dent* 2003; 6: 83–88.
- KURBAD A, REICHEL K: All-ceramic primary telescopic crowns with CEREC inLab. *Int J Comput Dent* 2003; 6: 103–11.
- MASEK R: Designing in 3D – a more visual approach to CEREC correlation. *Int J Comput Dent* 2003; 6: 75–82.
- REISS B: Occlusal surface design with CEREC 3D. *Int J Comput Dent* 2003; 6: 333–342.
- SCHENK O: CEREC Classic: The inlay with CEREC 3D step-by-step. *Int J Comput Dent* 2003; 6: 67–73.
- SCHNEIDER O: CEREC veneers – Practical procedure and case presentation. *Int J Comput Dent* 2003; 6: 283–292.
- SCHNEIDER W: CEREC 3D – a new dimension in treatment. *Int J Comput Dent* 2003; 6: 57–66.
- WIEDHAHN K: The CEREC 3D veneer program – is bad news good news? *Int J Comput Dent* 2003; 6: 95–101.
- 2004 FRITSCH G: Treatment of a single-tooth gap with CEREC 3D crown on an implant: A case report; *Int J Comput Dent* 2004; 7: 199–206.
- MACLEOD R: Chairside correlation in CEREC 3D software. *Int J Comput Dent* 2004; 7: 269–278.
- MASEK R: Margin isolation for optical impressions and adhesion. *Int J Comput Dent* 2005; 8: 69–76.
- MÖRMANN WH, BINDL A: 3D-CAD/CAM für jedermann in Praxis und Labor. *Zahnärztliche Mitteilungen* 2004; 94: 228–232.
- OTTO T: Computer-aided direct all-ceramic crowns: preliminary 1-year results of a prospective clinical study. *Int J Perio Rest Dent* 2004; 24: 447–455.
- REICH SM, WICHMANN M, RINNE H, SHORTALL A: Clinical performance of large, all-ceramic CAD/CAM-generated restorations after three years: a pilot study. *J Am Dent Assoc* 2004; 135: 605–612.
- WANG HR, ZHAO YF, ZHANG FM: Study on the occlusal surface design methods of CAD/CAM all-ceramic coping crown. *Sichuan Da Xue Xue Bao Yi Xue Ban*. 2004; 35: 280–282. Chinese
- 2005 BINDL A, MÖRMANN WH: Marginal and internal fit of all-ceramic CAD/CAM crown-copings und chamfer preparations. *J Oral Rehabil* 2005; 32: 441–447.
- BINDL A, RICHTER B, MÖRMANN WH: Survival of ceramic CAD/CAM crowns bonded to preparations with reduced macroretention geometry. *Int J Prosthodont* 2005; 18: 219–224.
- 2007 Masek R, Ultimate CEREC Creations – Comprehensive Single Visit Esthetic Dentistry. In Mörmann WH (ed.) State of the Art of CAD/CAM Restorations, 20 Years of CEREC, Berlin: Quintessence, 2006: 131–138.
- 10 | CEREC QUALITY GUIDELINES**
- 2001 PRÖBSTER L: Sind vollkeramische Kronen und Brücken wissenschaftlich anerkannt? Gemeinsame Stellungnahme von DGZMK und DGZPW. *Dtsch Zahnärztl Z* 2001; 56: 575–576.
- 2005 SCHWEIZERISCHE ZAHNÄRZTE GESELLSCHAFT SSO: Qualitätsleitlinien in der Zahnmedizin, Restaurative Zahnmedizin. *Schweiz Monatsschr Zahnmed* 2005; 115: 71–91.

# ALWAYS AT THE FOREFRONT OF INNOVATION!

As global innovation leader for dental equipment, we continuously invest in research and thus in the future of modern dentistry. By networking digital technologies with integrated solutions and optimizing the treatment workflow, we create improved treatment results, more comfort and safety for the patient as well as time and cost savings in everyday work. The combination of constant innovative power and globally growing sales and service structures makes Sirona the global market leader trusted by thousands of practices and labs around the world.

**Enjoy every day. With Sirona.**



#### **CAD/CAM systems**

From pioneer to new standard. For almost 30 years we have been developing digital dentistry and creating new possibilities for the future practice and lab.

#### **Imaging systems**

Best image quality with the lowest dose. More than 100 years of developing x-rays for the dental practice make us the number 1 innovation partner.

#### **Treatment centers**

The business card of modern practices. We are striving to create the ideal ergonomic and innovative center. Individually tailored to the well-being and demands of the patient and dentist.

#### **Instruments**

Advantages that speak for themselves. We make sure that we provide the right balance of proven quality, individual ergonomics and innovative technology for user-friendly work.

#### **Hygiene systems**

Competence that gives you safety. When it comes to hygiene in the practice, we do not take any shortcuts.